

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/29/2025

OPERATOR NAME:
MATTHEW C CADWELL
Permit #: 53014
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28

LEB

