

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/11/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

LEB

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:14
INTERNAL STANDARD	VERIFIED	21:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:14
BLANK TEST	0.00	21:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:16
BLANK TEST	0.00	21:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:17
BLANK TEST	0.00	21:18

