

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/18/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:46
INTERNAL STANDARD	VERIFIED	21:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:47
BLANK TEST	0.00	21:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:48
BLANK TEST	0.00	21:49

