

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/11/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:17
INTERNAL STANDARD	VERIFIED	23:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:18
BLANK TEST	0.00	23:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:20

LEB

