

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/16/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

LEB

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:56
INTERNAL STANDARD	VERIFIED	19:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:59
BLANK TEST	0.00	19:59

