

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/23/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:09
INTERNAL STANDARD	VERIFIED	20:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:10
BLANK TEST	0.00	20:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:12
BLANK TEST	0.00	20:13

LEB

