

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 02/19/2025

OPERATOR NAME:  
JASON L OSIKA ...  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 20:15 |
| INTERNAL STANDARD       | VERIFIED | 20:15 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 20:15 |
| BLANK TEST              | 0.00     | 20:16 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 20:17 |
| BLANK TEST              | 0.00     | 20:18 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 20:18 |
| BLANK TEST              | 0.00     | 20:19 |

