

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/05/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:07
INTERNAL STANDARD	VERIFIED	23:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:07
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:11

