

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:58
INTERNAL STANDARD	VERIFIED	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:01

