

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:07
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:10
BLANK TEST	0.00	02:11

