

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/01/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:54
INTERNAL STANDARD	VERIFIED	16:55
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	16:55
BLANK TEST	0.00	16:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	16:56
BLANK TEST	0.00	16:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	16:57
BLANK TEST	0.00	16:58

WEB

