

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 03/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:21
INTERNAL STANDARD	VERIFIED	00:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:23
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:25

