

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE

b3 010-45/PC#116 @ 360 ! L! α V H ↓ - ↓ -

Date: 04/01/2026

OPERATOR NAME: *Ux3*
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:11
INTERNAL STANDARD	VERIFIED	19:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:13
BLANK TEST	0.00	19:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:14
BLANK TEST	0.00	19:15

