

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/13/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:55
INTERNAL STANDARD	VERIFIED	14:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59

LEB

