

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/17/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO#
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:57
INTERNAL STANDARD	VERIFIED	20:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:58
BLANK TEST	0.00	20:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:59
BLANK TEST	0.00	21:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:00
BLANK TEST	0.00	21:01

LEB

