

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/16/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:45
INTERNAL STANDARD	VERIFIED	17:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49

