

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/16/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:13
INTERNAL STANDARD	VERIFIED	15:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:14
BLANK TEST	0.00	15:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:15
BLANK TEST	0.00	15:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:16
BLANK TEST	0.00	15:17

