

# SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 10/28/2026

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 26270

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 00:01 |
| INTERNAL STANDARD       | VERIFIED | 00:01 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 00:01 |
| BLANK TEST              | 0.00     | 00:02 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 00:03 |
| BLANK TEST              | 0.00     | 00:04 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 00:04 |
| BLANK TEST              | 0.00     | 00:05 |

LEB

