

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/16/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:17
INTERNAL STANDARD	VERIFIED	19:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:18
BLANK TEST	0.00	19:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:19
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:20
BLANK TEST	0.00	19:21

