

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/05/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:44
INTERNAL STANDARD	VERIFIED	18:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:44
BLANK TEST	0.00	18:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:45
BLANK TEST	0.00	18:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:47
BLANK TEST	0.00	18:47

