

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/11/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:18
INTERNAL STANDARD	VERIFIED	19:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:18
BLANK TEST	0.00	19:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:20
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:21
BLANK TEST	0.00	19:22

LEB

