

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/08/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:06
INTERNAL STANDARD	VERIFIED	19:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:09
BLANK TEST	0.00	19:10

