

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/04/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:01
INTERNAL STANDARD	VERIFIED	19:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:02
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:03
BLANK TEST	0.00	19:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:04
BLANK TEST	0.00	19:05

