

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/22/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26120

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:10
INTERNAL STANDARD	VERIFIED	03:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:10
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:12
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:13
BLANK TEST	0.00	03:14

