

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/26/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:41
INTERNAL STANDARD	VERIFIED	20:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:43
BLANK TEST	0.00	20:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:44
BLANK TEST	0.00	20:45

