

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/23/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:30
INTERNAL STANDARD	VERIFIED.	22:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:31
BLANK TEST	0.00	22:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:32
BLANK TEST	0.00	22:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34

LEB

