

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/09/2026

OPERATOR NAME:
STEVEN P EIELRS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:35
INTERNAL STANDARD	VERIFIED	15:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:36
BLANK TEST	0.00	15:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:37
BLANK TEST	0.00	15:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:38
BLANK TEST	0.00	15:39

