

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/28/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07

