

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/11/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:25
INTERNAL STANDARD	VERIFIED	20:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:25
BLANK TEST	0.00	20:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:27
BLANK TEST	0.00	20:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:28
BLANK TEST	0.00	20:29

