

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/04/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:21
INTERNAL STANDARD	VERIFIED	05:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:24
BLANK TEST	0.00	05:25

LEB

26

