

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 01/21/2026

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:00
INTERNAL STANDARD	VERIFIED	21:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:00
BLANK TEST	0.00	21:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:01
BLANK TEST	0.00	21:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:03
BLANK TEST	0.00	21:04

