

SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/07/2026

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:15
INTERNAL STANDARD	VERIFIED	05:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:15
BLANK TEST	0.00	05:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:17
BLANK TEST	0.00	05:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:18
BLANK TEST	0.00	05:19



LEB Bw #66