

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/25/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:24
INTERNAL STANDARD	VERIFIED	19:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:26
BLANK TEST	0.00	19:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:27
BLANK TEST	0.00	19:28

LEB

