

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/29/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:40
INTERNAL STANDARD	VERIFIED	19:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:40
BLANK TEST	0.00	19:41
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:42
BLANK TEST	0.00	19:43

