

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/13/2026

LLB

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:56
INTERNAL STANDARD	VERIFIED	18:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:57
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:58
BLANK TEST	0.00	18:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:59
BLANK TEST	0.00	19:00

