

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/01/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:27
INTERNAL STANDARD	VERIFIED	21:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:27
BLANK TEST	0.00	21:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:28
BLANK TEST	0.00	21:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:30
BLANK TEST	0.00	21:30

