

SUPERVISOR TEST




ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 08/11/2021

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37





SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 08/18/2021

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:21
INTERNAL STANDARD	VERIFIED	21:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:24
BLANK TEST	0.00	21:25

