

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/29/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:16
INTERNAL STANDARD	VERIFIED	17:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:17
BLANK TEST	0.00	17:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:18
BLANK TEST	0.00	17:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:19
BLANK TEST	0.00	17:20

