

SUPERVISOR TEST

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197 WFO

Date: 10/22/2021

OPERATOR NAME:
 ROBERT S CORNISH
 Permit #: 48068
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:06
INTERNAL STANDARD	VERIFIED	03:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10

