

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/11/2021

OPERATOR NAME:
JAMES P QUINN
Permit #: 49340
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:58
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:59
BLANK TEST	0.00	03:00

WFO Datamaster
weekly test
11/11/21

