

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/24/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:33
INTERNAL STANDARD	VERIFIED	20:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:34
BLANK TEST	0.00	20:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:35
BLANK TEST	0.00	20:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:36
BLANK TEST	0.00	20:37

