

SUPERVISOR TEST

w F D

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100197

Date: 12/15/2021

OPERATOR NAME:  
 BRETT A GRABELDINGER  
 Permit #: 46896  
 Operator Agency/Dept: ONEIDA CO. S.O.  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:26
INTERNAL STANDARD	VERIFIED	20:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:26
BLANK TEST	0.00	20:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:27
BLANK TEST	0.00	20:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:29
BLANK TEST	0.00	20:29

