

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/22/2021

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:22
INTERNAL STANDARD	VERIFIED	21:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:24
BLANK TEST	0.00	21:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:25
BLANK TEST	0.00	21:26

