

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/30/2021

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:50
INTERNAL STANDARD	VERIFIED	01:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:53
BLANK TEST	0.00	01:54

