

WFO

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/01/2022

OPERATOR NAME:
ROBERT S CORNISH
Permit #: 48068
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:11
INTERNAL STANDARD	VERIFIED	02:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:12
BLANK TEST	0.00	02:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:14
BLANK TEST	0.00	02:15

