

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

WFO

Date: 04/14/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:15
INTERNAL STANDARD	VERIFIED	06:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:17
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:18
BLANK TEST	0.00	06:19

