

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/28/2022

OPERATOR NAME:
JAY V SORRENTINO
Permit #: 38962
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:24
INTERNAL STANDARD	VERIFIED	02:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:27
BLANK TEST	0.00	02:28

