

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/09/2022

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:20
INTERNAL STANDARD	VERIFIED	05:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:21
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:22
BLANK TEST	0.00	05:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:23
BLANK TEST	0.00	05:24

