

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 03/09/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:08
INTERNAL STANDARD	VERIFIED	05:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:08
BLANK TEST	0.00	05:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:09
BLANK TEST	0.00	05:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:10
BLANK TEST	0.00	05:11

