

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 04/20/2022

WFO

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:43
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:44
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:45
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:46
BLANK TEST	0.00	02:47

